

PCT

WORLD INTELLECTUAL PROPERTY ORGANIZATION
International Bureau




INTERNATIONAL APPLICATION PUBLISHED UNDER THE PATENT COOPERATION TREATY (PCT)

(51) International Patent Classification ⁶ : A61B 5/00		A1	(11) International Publication Number: WO 99/39631 (43) International Publication Date: 12 August 1999 (12.08.99)
(21) International Application Number: PCT/US99/02586 (22) International Filing Date: 5 February 1999 (05.02.99) (30) Priority Data: 60/073,784 5 February 1998 (05.02.98) US (71) Applicant: IN-LINE DIAGNOSTICS CORPORATION [US/US]; Suite 160, 4646 South 1500 West, Riverdale, UT 84405 (US). (72) Inventors: STEUER, Robert, R.; 4157 North 900 West, Pleasant View, UT 84414 (US). MILLER, David, R.; 6551 North Trappers Loop, Morgan, UT 84050 (US). (74) Agents: MELSER, Allen, S. et al.; Thelen Reid & Priest LLP, Suite 800, 701 Pennsylvania Avenue, N.W., Washington, DC 20004 (US).			(81) Designated States: AL, AM, AT, AU, AZ, BA, BB, BG, BR, BY, CA, CH, CN, CU, CZ, DE, DK, EE, ES, FI, GB, GE, GH, HR, HU, ID, IL, IN, IS, JP, KE, KG, KP, KR, KZ, LC, LK, LR, LS, LT, LU, LV, MD, MG, MK, MN, MW, MX, NO, NZ, PL, PT, RO, RU, SD, SE, SG, SI, SK, SL, TJ, TM, TR, TT, UA, UG, UZ, VN, YU, ZW, ARIPO patent (GH, GM, KE, LS, MW, SD, SZ, UG, ZW), Eurasian patent (AM, AZ, BY, KG, KZ, MD, RU, TJ, TM), European patent (AT, BE, CH, CY, DE, DK, ES, FI, FR, GB, GR, IE, IT, LU, MC, NL, PT, SE), OAPI patent (BF, BJ, CF, CG, CI, CM, GA, GN, GW, ML, MR, NE, SN, TD, TG). Published With international search report. EP 1052930

(54) Title: METHOD AND APPARATUS FOR NON-INVASIVE BLOOD CONSTITUENT MONITORING

(57) Abstract

This invention is a system for determining a biologic constituent including hematocrit trans-cutaneously, non-invasive, and continuously. A finger clip assembly (6) includes at least a pair of emitters, and a photo-diode in appropriate alignment to enable operation in either a transmissive mode or a reflectance mode. At least one predetermined wavelength of light is passed into or through body tissues (7) such as a finger, earlobe, or scalp, etc., and attenuation of light at that wavelength is detected. Likewise, the change in blood flow is determined by various techniques including optical, pressure, piezo and strain gage methods. Mathematical manipulation of the detected values compensates for the effects of body tissue, fluid, and determines the hematocrit value. If an additional wavelength of light is used which attenuates light substantially differently by oxyhemoglobin, and reduced hemoglobin, then the blood oxygen saturation value, independent of hematocrit may be determined. Further, if an additional wavelength of light is used which greatly attenuates light due to bilirubin (440 nm) or glucose (1060 nm), then the bilirubin or glucose value may also be determined. Also, determination of the hematocrit with a two step DC analysis technique is disclosed.



FOR THE PURPOSES OF INFORMATION ONLY

Codes used to identify States party to the PCT on the front pages of pamphlets publishing international applications under the PCT.

AL	Albania	ES	Spain	LS	Lesotho	SI	Slovenia
AM	Armenia	FI	Finland	LT	Lithuania	SK	Slovakia
AT	Austria	FR	France	LU	Luxembourg	SN	Senegal
AU	Australia	GA	Gabon	LV	Latvia	SZ	Swaziland
AZ	Azerbaijan	GB	United Kingdom	MC	Monaco	TD	Chad
BA	Bosnia and Herzegovina	GE	Georgia	MD	Republic of Moldova	TG	Togo
BB	Barbados	GH	Ghana	MG	Madagascar	TJ	Tajikistan
BE	Belgium	GN	Guinea	MK	The former Yugoslav	TM	Turkmenistan
BF	Burkina Faso	GR	Greece		Republic of Macedonia	TR	Turkey
BG	Bulgaria	HU	Hungary	ML	Mali	TT	Trinidad and Tobago
BJ	Benin	IE	Ireland	MN	Mongolia	UA	Ukraine
BR	Brazil	IL	Israel	MR	Mauritania	UG	Uganda
DY	Belarus	IS	Iceland	MW	Malawi	US	United States of America
CA	Canada	IT	Italy	MX	Mexico	UZ	Uzbekistan
CF	Central African Republic	JP	Japan	NE	Niger	VN	Viet Nam
CG	Congo	KE	Kenya	NL	Netherlands	YU	Yugoslavia
CH	Switzerland	KG	Kyrgyzstan	NO	Norway	ZW	Zimbabwe
CI	Côte d'Ivoire	KP	Democratic People's	NZ	New Zealand		
CM	Cameroon		Republic of Korea	PL	Poland		
CN	China	KR	Republic of Korea	PT	Portugal		
CU	Cuba	KZ	Kazakhstan	RO	Romania		
CZ	Czech Republic	LC	Saint Lucia	RU	Russian Federation		
DE	Germany	LI	Liechtenstein	SD	Sudan		
DK	Denmark	LK	Sri Lanka	SE	Sweden		
EE	Estonia	LR	Liberia	SG	Singapore		

METHOD AND APPARATUS FOR NON-INVASIVE BLOOD CONSTITUENT MONITORING

BACKGROUND

The present invention is related to U.S. Patent Numbers 5,372,136 and 5,499,627 the text and drawings of which are incorporated herein by reference as if reproduced in full.

THE FIELD OF INVENTION

The present invention relates to improvements in the systems and methods for non-invasively measuring one or more biologic constituent concentration values. More particularly, the present invention relates to non-invasive spectrophotometric systems and methods for quantitatively and continuously monitoring the hematocrit and other blood parameters.

THE PRIOR ART

Modern medical practice utilizes a number of procedures and indicators to assess a patient's condition. One of these indicators is the patient's hematocrit. Hematocrit (often abbreviated as HCT) is the volume expressed as a percentage of the patient's blood which is occupied by red corpuscles, commonly referred to as red blood cells. The present invention is presented in the context of hematocrit. However, it is to be understood that the teachings of the present invention apply to any desired biologic constituent parameter.

Medical professionals routinely desire to know the hematocrit of a patient. In order to determine hematocrit using any of the techniques available to date, it is necessary to draw a sample of blood by puncturing a vein or invading a capillary. Then, using widely accepted techniques, the sample of blood is subjected to either high-speed centrifuge, cell counting, ultrasonic, conductometric or photometric methods of evaluating the sample of blood in a fixed

container. Prior U.S. Patent Number 5,372,136 indicates a system and methodology for determining the hematocrit non-invasively, without puncturing or invading the body, spectrophotometrically and continuously in a subject. The present invention relates to improvements upon the above cited system.

Beyond the above referenced patent, others have suggested various means of noninvasive measurement of hematocrit. Specifically, Mendelson, U.S. Patent Number 5,277,181; Seeker, U.S. Patent 5,188,108; Gonatas, U.S. Patent Number 5,528,365; Ishikawa, U.S. Patent Number 5,522,388; Shiga, U.S. Patent Number 4,927,264; Tsuchiya, U.S. Patent Numbers 5,441,054, 5,529,065, 5,517,987 and 5,477,051; and Chance, U.S. Patent Numbers 5,353,799, 5,402,778, and 5,673,701 have attempted to define means of directly measuring desired biologic constituents such as hematocrit. Even though the various patents indicate the need to utilize multiple wavelengths measured at different detection sites and/or the need to perform differential or ratiometric operations on the detected optical signal, all fail to isolate and resolve the individual and specific scattering and absorption coefficients of the desired constituent. At best they address only bulk attenuation coefficients and/or bulk diffusion constants of the scattering media while attempting to resolve such constraints as tissue nonhomogeneity. As an example, tissue may be considered to contain a bulk absorptive coefficient due to blood, collagen, water, fibers, bone, fingernail, etc. Hence, in order to determine the absorptive coefficient of the blood itself, the bulk value of the tissue per se must be prorated by the amounts of the above constituents. Secondly, the actual absorptive coefficient of the blood must then be decoupled or isolated from its proration factor as well.

OBJECTS OF THE INVENTION

Thus, it is an object of the present invention to provide an improvement in the systems and methods for the non-invasive (transcutaneous) and continuous determination of the blood Hematocrit in living tissue.

It is yet another object of the present invention to provide an improvement in the systems and methods for the non-invasive (transcutaneous) and continuous determination of the blood constituents, including glucose, bilirubin, cholesterol, tissue water, etc. in living tissue.

It is another object of the present invention to provide a system and method and apparatus for the display of both immediate and/or continuous visual information regarding the HCT of the subject.

It is yet another object of the present invention to provide a repeatable and reliable method and apparatus for the non-invasive determination of hematocrit transcutaneously and in real time even under varying physiological conditions.

Still another object of the present invention is to provide a method and apparatus for the instantaneous determination of the bulk absorption coefficient of the scattering media.

These and other objects and advantages of the invention will become more fully apparent from the description in the specification and claims, which follow.

SUMMARY OF THE INVENTION

In one aspect, the present invention accomplishes the transcutaneous, noninvasive, real-time and continuous measurement of the hematocrit and other blood constituents of the patient. That is, the electronic circuitry necessary is included to receive signals from a detector and to generate appropriate signals at various input sites as described in U.S. Patent Number 5,372,136.

Yet another aspect of the present invention is the ability to extract the blood absorption coefficient from the bulk tissue diffusion constant or the bulk absorption coefficient of the scattering media by requiring both physical and mathematical operations.

BRIEF DESCRIPTION OF THE DRAWINGS

Figures 1 and 1A show a finger placed into a clam-shell type fixture constituting a receiving means for detector and emitter arrays operating in a transmission mode and the blood conduit which in the figures is the finger.

Figures 1B and 1C are similar to Figure 1A, but show the detector and emitter arrays operating in a reflectance mode.

Figure 1D is a schematic diagram for a mylar base with a detector, emitters and either a strain gage or a pressure transducer for inclusion in the clam-shell fixture.

Figure 1E is a schematic diagram for a detector emitter array using a single, moveable emitter in a transmission mode.

Figure 2 shows actual patient data plot of $\ln(i)$ vs. d .

Figure 3 illustrates actual patient data of the $(\partial i / \partial t) / i$ dependence on d .

Figure 4 shows the blood absorption coefficient's dependence on hematocrit.

Figure 5 indicates the nonlinear relationship between the V_e / V_f and pressure.

Figure 6 shows the electrical circuit diagram of the piezo film / strain gage transducer means.

Figure 7 is the plot of $f(H)$ vs. measured Hematocrit.

Figure 8 shows the instantaneous time derivatives of $(\partial i / \partial t) / i$ and $\partial P / \partial t$ versus time during one cardiac pulse.

Figure 9 plots $(\partial i / \partial t) / i$ versus $\partial P / \partial t$ for a given human pulse at d_1 , d_2 , d_3 , and d_4 .

Figure 10 plots $(\partial i / \partial t) / (\partial P / \partial t)$ versus time during a single cardiac pulse cycle.

Figure 11 is the circuit diagram of the pressure transducer means.

Figure 12 plots α versus X_0 at a fixed Hematocrit.

Figure 13 gives the patient data of the new transcutaneous Hematocrit method and system plotted versus the measured Hematocrit standard.

DETAILED DESCRIPTION OF THE INVENTION

In a preferred embodiment of the invention, measurements are conducted using a modified version of the apparatus described in U.S. Patent Numbers 5,456,253 and 5,372,136, both of which are incorporated herein as if reproduced in full below. Both of these patents form part of the present disclosure.

Thus, in a preferred embodiment, hematocrit is measured in living tissue located at some convenient location on the body, such as, an ear lobe, finger tip, nose or other accessible tissue sites. In a preferred embodiment the apparatus and signal manipulations described in U.S. Patent Number 5,372,136 are utilized to measure various optical parameters that will be described hereafter. The numbered components in Figures 1, 1A, 1B, and 1C are similar to the numbers in Figure 1 of U.S. Patent Number 5,456,253.

In the present disclosure, Figure 1 shows the finger 7 of an individual placed into a clam-shell type fixture 6 wherein the optical and other physical measurements can be easily performed. The clam-shell type holder allows for adaptability to various finger sizes. However, other fixture methods such as Figures 1B through 1E, can be used to obtain similar physical data as using the clam-shell fixture.

THEORETICAL BASIS OF THE SPECTROPHOTOMETRIC AND MATHEMATICAL ANALYSIS FOR TRANSCUTANEOUS HEMATOCRIT MEASUREMENT

Non-invasive, transcutaneous hematocrit measurement using a spectroscopic method is described below:

I. Introduction

Earlier spectrophotometric techniques have fallen short of being able to fully characterize the individual blood absorbance coefficients. The following discussion demonstrates the method of decoupling, or isolating from the bulk tissue attenuation parameters (including the convoluted absorptive and scattering parameters) the individual blood absorptive constants. This unique method identifies, isolates and compartmentalizes each of the contributing biologic elements of the tissue media. This decoupling process can either isolate the blood absorbance of interest and/or eliminate the scattering contribution from the bulk media measurement.

From photon diffusion analysis:

$$\frac{\partial^2 \Psi(\rho)}{\partial \rho^2} - \alpha^2 \Psi(\rho) = -\frac{S(\rho)}{D} \quad (1)$$

where,

$$D = \frac{1}{3(K + S)} \quad (2)$$

$$\alpha = \sqrt{3K(K + S)} \quad (3)$$

$$K = K_b X_b + K_s X_s + K_w X_w \quad (4)$$

$$K_b = \frac{H}{V} (\sigma_{a,r} SAT + (1 - SAT)\sigma_{a,r}) + (1 - H)K_p \quad (5)$$

$$S = S_b X_b + S_s X_s \quad (6)$$

$$S_b = \frac{\sigma_a H (1-H) (1.4-H)}{V} \quad (7)$$

and where,

α = Bulk attenuation coefficient of the tissue sample

K = Bulk absorption coefficient of the tissue sample

S = Bulk scattering coefficient of the tissue sample

D = Diffusion constant

K_b = Macroscopic absorption coefficient for whole blood (WB)

S_b = Macroscopic transport - corrected scattering coefficient for WB

K_p = Macroscopic absorption coefficient for plasma

K_s = Macroscopic absorption coefficient for skin, & other non water/blood components

K_w = Macroscopic absorption coefficient for water

V = Volume of a red blood cell (RBC)

H = Hematocrit, volume fraction of RBCs to total blood volume

SAT = Oxygen saturation %

σ_{ao} = Absorption cross - section of oxygenated RBCs

σ_{or} = Absorption cross - section of deoxygenated RBCs

σ_s = Transport - corrected scattering cross-section of RBCs

X_b = Fractional volume of blood per total tissue volume

X_s = Fractional volume of skin, & non water/blood components per total tissue volume

Alternatively from (13a) : $H = \frac{\alpha \cdot (\partial I / \partial t / I)}{(d - d_0) \cdot X'_b}$ (18b)

Equation (18a) indicates a small nonlinearity in H may occur based on the magnitude of K for a given individual.

It should be reiterated that the change in received intensity with time is a result of the change in normalized blood volume resulting from the cardiac cycle itself as blood pulses through the examined tissue. As the intensity of the received light is measured, its time rate of change can be calculated. The change with distance can be determined by placing multiple emitters (such as 1-4 in Figure 1 A) and/or multiple detectors such that multiple thicknesses of tissue and hence, lengths of tissue are penetrated.

To examine $\frac{\partial X_b}{\partial t}$ further, the following can be defined for the illuminated tissue:

V_b = Volume of blood,

V_w = Volume of water, and

V_s = Volume of skin, tissue and other non-water or blood components.

By definition,

$$X_b = \frac{V_b}{V_b + V_w + V_s} \quad (20)$$

differentiating (20) with respect to time gives:

$$\frac{\partial X_b}{\partial t} = \frac{(V_w + V_s) \frac{\partial V_b}{\partial t} - V_b \frac{\partial V_w}{\partial t}}{(V_b + V_w + V_s)^2} \quad (21)$$

Since $\frac{\partial V_w}{\partial t} \ll \frac{\partial V_b}{\partial t}$ and $V_b \ll V_w + V_s$, (21) simplifies to:

$$\frac{\partial X_b}{\partial t} = \frac{\frac{\partial V_b}{\partial t}}{V_{total}} \quad (22)$$

It is emphasized that α is a function of the bulk absorption and scattering coefficients, K and S , as well as hematocrit, H . Further, that K and S are functions of the fractional volumes of each constituent, X_b , X_r , and X_w , which must be used to prorate the individual absorption and scattering coefficients, K_b , K_r , K_w , S_b and S_r . Therefore, the transducer system must be responsive not only to a change in volume (ΔV) due to the influx of the blood, but must also be responsive to the normalized change in volume of blood, normalized to the total volume of the finger (V_f) or tissue being measured,

$$\left(\frac{\Delta V_f}{V_f} \right).$$

For Reflectance (R) measurements in homogenous tissue:

$$R = Ae^{-\alpha r}, \text{ where } A = (1/r^2 + 1/\alpha r), \text{ where } r \text{ is the radial distance,}$$

and

$$\frac{dR}{R} = \alpha' (r/(\alpha^2 + \alpha r))$$

However, for tissue, which is typically non-homogeneous with a dermal and subcutaneous layer, the reflectance will not be a trivial function but can be described as approximately:

$$R = [(C_1 + C_2) \exp(-C_3 \cdot r)]/r^n$$

Where C_1 and C_2 are inter-related photon flux densities between the dermal layer 12 and the subcutaneous layer, 12a (see Figures 1C and 1E). Likewise, C_3 is a strong function of z_1 , z_2 ,

α_1 , and α_2 ; i.e., the thickness of the dermis or dermal layer 12, subcutaneous layer 12a, and their respective α 's.

Since C_3' is a function of the inter-related photon flux densities C_1 and/or C_2 and if Xb_1' does not equal Xb_2' , then the slope C_3' will not be nulled out by the Xb' monitors mentioned. Therefore, Xb_2' must be greater than Xb_1' . Then the pressure or piezo monitors will compensate correctly. The circular pressure balloon is ideal for not only sensing the change in a pressure, but also providing a pressure against the dermis causing Xb_1' to be small. However, recognizing that the penetration depth of the 800 nm light typically extends through dermal layer 12 into the deep tissue, subcutaneous layer 12a, a different wavelength selection is appropriate. Thusly, when the photons only penetrate into the dermal layer 12, C_3' will only be a function of z_1 and α_1 . Those selected wavelengths, as mentioned in U.S. Patent No. 5,372,136, would be the green (570-595 nm) wavelength and 1300 nm wavelength. The green wavelengths are used as the hematocrit bearing wavelength and the 1300 nm wavelength is used as the non-hematocrit bearing, or reference wavelength. That is, for reflectance measurements the green (Gr)-1300 wavelength pair would give the hematocrit information as:

$$\frac{\Delta Gr / Gr}{\Delta 1300 / 1300} \cdot \frac{\alpha_{Gr}}{\alpha_{1300}} = f(HCT)$$

III. Methods of $\frac{\partial X}{\partial t}$ measurement

$\partial X / \partial t$ can be measured and compensated for through the use of a number of different methods - (a) a pressure transducer, (b) a strain transducer such as piezo electric film or

strain gage, (e) a different wavelength of light, such as 1300 nm, which also holds $\partial X_f / \partial t$ information, but holds little hematocrit information, or (d) other transducers. The individual methods of obtaining $\partial X_f / \partial t$ are addressed below.

A. Pressure Transducer Measurement of $\frac{\partial X_f}{\partial t}$

Consider a pressure transducer system 36 with a gas filled bladder 38 surrounding a finger tip 10 of a patient contained within a fixed volume clam shell fixture 6, see Figures 1, 1A- 1D. The same derivations, equations, and results would apply to any other body appendage or tissue that could be contacted such that a change in the tissue volume would change the pressure of the contacted pressure transducer system. For a finger note:

$$V_{clam} = V_{sys} + V_f \quad (23)$$

where

V_{clam} = Clam-shell fixture volume

V_{sys} = Bladder system volume

V_f = Finger volume

Also $\Delta V_f = -\Delta V_{sys}$. The system will have a bulk modulus of elasticity, β , such that:

$$\frac{\Delta V_{sys}}{V_{sys}} = -\frac{\Delta P_{sys}}{\beta} = -\frac{\Delta V_f}{V_{sys}} \quad (24)$$

Substituting (23) into (24) results in:

$$\frac{\Delta V_f}{V_f} = \left(\frac{V_{clam}}{V_f} - 1 \right) \frac{\Delta P_{sys}}{\beta} \quad (25)$$

Since $\Delta V_f = \Delta V_b$ then from (25) we have:

$$\frac{\partial X_b}{\partial t} = \left(\frac{V_{clam}}{V_f} - 1 \right) \frac{\Delta P_{sys}}{\beta} \quad (25a)$$

As stated above, β is a constant of the pressure transducer system. However, an empirical solution for $\left(\frac{V_{clam}}{V_f} - 1 \right)$ was found to have a nonlinear relation to the pressure of the transducer system. For a given clam shell - pressure transducer embodiment a polynomial, $F(p)$, can accurately describe $\left(\frac{V_{clam}}{V_f} - 1 \right)$, see Figure 5

B. Strain Transducer (Strain Gage/Piezo Electric Film) Measurement of $\frac{\partial X_b}{\partial t}$

Again it is assumed that $\Delta V_b = \Delta V_f$ and that the finger changes volume only by a change in diameter. A strain gage or piezo electric film is secured tightly around the finger (again any applicable body appendage or tissue would apply) such that a change in diameter would produce a strain in the transducer. Specifically assuming a cylindrical finger:

$$\frac{\partial V_b}{\partial t} = \frac{\partial V_f}{\partial t} = \frac{\partial (\pi z r^2)}{\partial t} = 2\pi z r \frac{\partial r}{\partial t} \quad (26)$$

Normalizing with respect to V_f yields:

$$\frac{\partial X_b}{\partial t} = \frac{\frac{\partial V_b}{\partial t}}{V_{total}} = \frac{2\pi z r \frac{\partial r}{\partial t}}{\pi z r^2} = \frac{2}{r} \frac{\partial r}{\partial t} \quad (27)$$

A change in the length of the transducer element is related to a change in finger

radius by $\Delta L = 2\pi\Delta r$, therefore:

$$\frac{\partial X_b}{\partial t} = 2 \frac{(\partial L / \partial t)}{L_s} = 2 \gamma(t) \quad (28)$$

where $\gamma(t) = \frac{\partial L / \partial t}{L}$ is the rate of change in the strain as a function of time. For a strain

gage this value can be measured from an appropriate electrical circuit, see Figure 6, as it is proportional to the rate of change in the gage resistance.

For a piezo electric film the voltage produced is proportional to the strain, therefore:

$$\frac{\partial X_b}{\partial t} = \frac{2}{g_{31}\tau} \frac{\partial v(t)}{\partial t} \quad (29)$$

where, g_{31} is the piezoelectric coefficient for the stretch axis, τ is the film thickness and $v(t)$ is the open-circuit output voltage.

C. 1300 nm Light Measurement of $\frac{\partial X_b}{\partial t}$

The selection of the 1300 nm wavelength is based on criteria established in U.S.

Patent Number 5,372,136. The approach here is not to solve for $\partial X_b / \partial t$, and substitute into (19) but to ratiometrically eliminate $\partial X_b / \partial t$. In the case of the 1300 nm reference wavelength, the assumptions following equation (12) are no longer valid; i.e., $\partial X_b / \partial t$ and $\partial X_s / \partial t$ are not negligible, since water absorption at 1300 nm is so large. Hence, for the 1300 nm equations (13), (14) and (15) would result in:

$$\left(\frac{\partial \alpha}{\partial t} \right)_{11} = \frac{3}{2\alpha} \left[\{ (2K + S)K_b + KS_b \} \frac{\partial X_b}{\partial t} + \{ (2K + S)K_s + KS_s \} \frac{\partial X_s}{\partial t} + \right. \quad (30)$$

$$\left. \{ (2K + S)K_w \} \frac{\partial X_w}{\partial t} \right]$$

where, α , and the bulk and material specific K , and S are wavelength (λ) dependent.

Recalling that, $X_b + X_s + X_w = 1$, by definition, and that:

$$-\frac{\partial X_b}{\partial t} - \frac{\partial X_s}{\partial t} = \frac{\partial X_w}{\partial t} \quad (31)$$

By substituting (31) into (30) and noting that $K_{w,13} = K_{b,13}$, the following is obtained:

$$\left(\frac{\partial \alpha}{\partial t}\right)_{13} = \frac{3}{2\alpha} \left[\{KS_b\} \frac{\partial X_b}{\partial t} + \{(2K + S)(K_s - K_w) + K_s S_s\} \frac{\partial X_s}{\partial t} \right] \quad (32)$$

Since, $\frac{\partial X_b}{\partial t} \gg \frac{\partial X_s}{\partial t}$, (32) becomes:

$$\left(\frac{\partial \alpha}{\partial t}\right)_{13} = \frac{3}{2\alpha_{13}} \{KS_b\}_{13} \frac{\partial X_b}{\partial t} \quad (33)$$

Therefore, to eliminate $\frac{\partial X_b}{\partial t}$ and solve for the hematocrit, (17) is divided by (33)

yielding:

$$\frac{(\partial \alpha / \partial t)_s}{(\partial \alpha / \partial t)_{13}} = \frac{\alpha_{13}}{\alpha_s} \frac{K_{b,s} S_s}{K_{13} S_{b,13}} \quad (34)$$

Since S_s and K_{13} are well behaved and known (let $K_{13}/S_s = G$) in human tissue and the ratio $\frac{K_{b,s}}{S_{b,13}}$ is a function of H , then rearranging (34) gives:

$$f(H) = \frac{K_{b,s}}{S_{b,13}} = \frac{\alpha_s}{\alpha_{13}} \frac{\left(\frac{\partial \alpha}{\partial t}\right)_s}{\left(\frac{\partial \alpha}{\partial t}\right)_{13}} G \quad (35)$$

Where $\frac{\partial \alpha}{\partial t}$ can be measured using (11) or (14). See Figure 7 for $f(H)$.

D. Other $\partial X_y / \partial t$ measurements such as doppler, ultrasonic, electrical conductivity, magnetic permeability and other techniques have similar derivations. The important consideration is that $\partial X_y / \partial t$ is a normalized time varying quantity.

IV. Analytical implementation

If hematocrit is constant over a given time interval, averaging can eliminate system noise whose frequency components have corresponding periods much shorter than the interval. In addition, by observing the data variance during the interval it may be determined that the data is invalid. In the present system, the data acquisition rate is approximately 1000 data samples per second. This means that within a typical human pulse about 1000 samples of data are available for appropriate numerical analysis, averaging and qualification. Recognizing that both the intensity of light and the pressure in the transducer system are changing in time during the influx of blood is of great importance. Since the parametric relationship of $\partial \alpha / \partial t$ as a function of $\partial P / \partial t$ (where P is pressure) during the cardiac cycle should be linear, a multiplicity of data points facilitate qualification of the signal for accuracy and linearity. Whereas, prior techniques involving only the peak and valley values of the cardiac cycle require numerous pulses to qualify the data set. See Figures 8, 9 and 10.

Fig. 8 shows $di/dt/i$ as well as dP/dt verses time during the cardiac pulse - it is a pulse showing $\approx 200+$ data samples during the pulse.

Fig. 9 shows $(di/dt)/i$ vs dP/dt showing that within one cardiac pulse 200 plus data samples

are linearly related, i.e. trace up out of the "0" origin up to a maximum value and then back down toward the origin again.

Fig. 10 shows $d\alpha/dt/dP/dt$ versus time during one single cardiac pulse with 200 plus samples of data from time 15 - 45 giving a value of about 4.5 thousandths. The data can then be averaged, as if 200+ individuals pulse (max-min) values were actually taken as present day oxymeters do.

A. Homogeneity

Since the above derivations are based on the assumption of tissue homogeneity (i.e., $\partial X_b/\partial t = \partial X_b/\partial t$, $A_1 = A_2$, $\partial A_1/\partial X_b = \partial A_2/\partial X_b$, $\alpha_1 = \alpha_2$, etc.), high-speed, single-pulse, multiple parameter sampling allows for mathematical qualification of homogeneity, by requiring linearity of $\ln(i)$ vs. d and $(\partial i/\partial t)/i$ vs. d . Under these constraints and when qualified as homogeneous, $(\partial\alpha/\partial t)/(\partial P/\partial t)$ also may be assumed to be linear over the entire pulse contour. Finally, both α and $\partial\alpha/\partial t$ must also be linear, further assuring homogeneity in X_b and in $\partial X_b/\partial t$.

B. Circuitry

See U.S. Patent Number 5,372,136 for the operational circuitry description, which allows for high speed sampling of the optical intensities. See Figures 6 and 10 for similar circuitry considerations for sampling of pressure, peizo, and strain-gage measurements.

The circuitry shown and discussed in US Patent Number 5,372,136 is programmable by conventional techniques to solve and implement the equations and calculations presented in this application. Figure 6 shows a piezo transducer circuit having a transducer 50 connected to a series of operational amplifiers, resistors and capacitors in accordance with the figure. The circuit terminates in an analog output 52 for connection to the "E" connection shown in the middle left

side of Fig. 9D in U.S. Patent Number 5,372,136. Figure 11, on the other hand, shows a pressure transducer circuit having a pressure transducer made 62 connected to a series of operational amplifiers, a capacitor, resistors and variable resistors as shown in the figure. The circuit terminates in an analog output also connected to the aforementioned "E" connection.

Referring more specifically to FIG. 6, a crystal oscillator is connected to ground and to the non-inverting input of a first operational amplifier, which may be an LM158. The non-inverting input of the first operational amplifier is connected to ground by a .047 μ F capacitor C3. The first operational amplifier's feedback path to its inverting input includes a 470 K resistor R8. The first operational amplifier is suitably biased at the junction of a 220 Ω resistor R7 and a 150 μ F capacitor C4 that are connected between VCC and ground.

A second operational amplifier, which may also be an LM 158, receives the output of the first operational amplifier at its inverting input via a 10 K Ω resistor R5. The second operational amplifier's non-inverting input is connected to several locations:

- to a voltage VB51, which may be 4.096 volts, through a 10 K Ω resistor R2;
- to a middle node of a voltage divider, the voltage divider extending between the non-inverting input of the first operational amplifier via a 10 M Ω resistor R4 to the middle node, and via a 10 K Ω resistor R1 to ground;
- to the inverting input of the first operational amplifier via a 10 K Ω resistor R9; and
- to ground via a 220 μ F capacitor C5.

The second operational amplifier's feedback path to its inverting input includes a parallel arrangement of a 0.1 μ F capacitor C2 and a 47 K Ω resistor R6. The second operational amplifier drives the A/D output 52 via a 10 K Ω resistor R3, the output connected to ground via a 1 μ F

capacitor C1.

Of course, the particular choice, arrangement and values of components shown in FIG. 6 may be varied while still remaining within the scope of the invention.

Referring now to FIG. 10, first through fourth operational amplifiers, which may be LM348s, are illustrated. The operational amplifiers are powered and biased by voltages VCC and VEE.

The first operational amplifier's non-inverting input is set to a value determined by the tap setting of a 1 K Ω adjustable resistor R2 that extends between VCC and VEE. The DAC input drives the first operational amplifier's inverting input via a 1 K Ω resistor R1. The first operational amplifier's feedback path includes a 50 K Ω adjustable resistor R4. The first operational amplifier drives the second operational amplifier's inverting input through an 11 K Ω resistor R3. The feedback path to the inverting input of the second operational amplifier includes a 100 Ω resistor R5.

A transducer 62, which may include a Motorola MPX20100P, has opposite terminals that drive the non-inverting inputs of the second and third operational amplifiers, respectively. The other two opposite terminals of the transducer are connected to VCC and ground, respectively.

The second operational amplifier drives the inverting input of the third operational amplifier via a 750 Ω resistor R6. The third operational amplifier's feedback path to its inverting input includes a parallel arrangement of a 93.1 K Ω resistor R10 and a .001 μ F capacitor C1.

The third operational amplifier drives the non-inverting input of the fourth operational amplifier via a 1 K Ω resistor R7. The inverting input of the fourth operational amplifier is connected to ground via a 1 K Ω resistor R8. The feedback path to the inverting input of the

fourth operational amplifier includes a 50 K Ω adjustable resistor R9. The fourth operational amplifier drives the output of the FIG. 11 circuit.

Of course, the particular choice, arrangement and values of components shown in FIG. 10 may be varied while still remaining within the scope of the invention.

C. Preferred Embodiment

Physical embodiments as shown in Figure 1 include the optical array, pressure transducer/balloon system and clam-shell fixture. Requisites of the preferred embodiment include a holder for the finger (or other tissue) such as seen in Figures 1 and 1A and 1B. This clam-shell fixture not only secures the tissue but also the optical array, and transducer system.

Figure 1D is a schematic diagram for a mylar base member 38 that is shaped generally like a cross. As oriented in Figure 1D, vertically extending portion 52 crosses with a horizontally extending portion 54 to yield top leg 56, bottom leg 58, and side legs 60, 62. In use, a finger 7 lies along the longitudinally extending portion 52 with the finger tip placed on the top leg 56 to properly cover the arrangement of LED's 32 and photodetector 34, which are arranged like those on Figures 1A - 1C. A piezoelectric pressure transducer or strain gage 66 spans the horizontally extending portion 54 from near the tip of side leg 60 to the tip of side leg 62. In this orientation, the transducer or gage may be wrapped around the finger 7 for use in measurements.

The optical array 30, seen in Figure 1D, shows the arrangement of multiple LED's 32 spaced at known separation distances from the detector 34. This array provides for the instantaneous distance, or " d ", derivative, by the transmission mode shown in Figure 1A or in reflectance modes shown in Figures 1B and 1C. However, as shown in Figure 1E, a single LED 42 swept across the finger 7 or tissue surface 9 with a stepper motor 44 would provide a d

derivative as would a cantilevered clam-shell with an angular measurement device. In any case, d must be known and/or fixed. Also the detectors and emitters may be placed anywhere about the finger.

The pressure/balloon, strain gage, or peizo transducer system incorporated within the clam-shell fixture (see Section III, A, B, C and Figure 1A) provides the contact surface area needed to define the $\partial X_i/\partial t$.

High-speed sampling provides for a closer approximation of the instantaneous time, t , derivative, $\partial/\partial t$, as opposed to peak-valley values, see Figure 8. Therefore, the above embodiments allow for the direct measurement of $\ln(i)$ at d_1 , d_2 , d_3 and d_4 cotemporaneously, thereby determining the actual α of the sampled tissue. Likewise $(\partial i/\partial t)/i$ can be directly measured at d_1 , d_2 , d_3 and d_4 cotemporaneously during the pulse which determines the instantaneous $\partial\alpha/\partial t$.

The above mentioned optical array can be utilized transmissively and/or reflectively provided the separation distance between the detector and first emitter (d_1) is greater than 3mm.

D. Choice of Non-Ionizing Wavelengths

Since hematocrit is an example of the desired biological constituent concentration value of interest, selection criteria of the preferred wavelength must include an understanding of equation (5). That is, a wavelength whose coefficients K_s , K_w , K_p are small compared to K_h and which are also insensitive to oxygen saturation status must be selected. Such wavelengths include 805 nm, 590 nm, 569 nm and other isobestic wavelengths with negligible water absorption. While non-isobestic wavelengths, with small water absorption, could function, a second wavelength is needed to null out the oxygen saturation effects.

If the desired biologic constituent value of interest is the blood glucose, bilirubin, cholesterol or other parameters, then a second wavelength must be chosen. The first wavelength, 805 nm, is used to measure the hematocrit, H , after which a K_{p805} (the absorbance of plasma at $\lambda = 805$ nm) can be determined. Then, knowing the H , a second wavelength, 570 nm, is chosen where K_{p570} is less than K_{p805} . Similarly, if the first wavelength used to measure the H and the reference glucose, K_p (glucose) is 570 nm, the second wavelength, 1060 nm, is chosen where K_{p570} is much less than K_{p1060} . In the case of bilirubin, the first wavelength used to measure the H and the reference bilirubin, K_p (bilirubin), is 570 nm, the second wavelength, 440 nm, is then chosen when K_{p570} is much less than K_{p440} . The selection of these above mentioned wavelengths therefore assures uniqueness for the measurement of the desired biologic constituent.

Additionally for glucose determination, recall that the 1300 nm wavelength is not hematocrit or hemoglobin dependent but will be glucose sensitive. This is primarily due to the dependence of the scattering coefficient on the difference between the index of refraction of pure water and glucose, i.e.: recall

$$S_{b8} = H (1-H) \sigma_{s8} \text{ (from equation 7) where:}$$

$$\sigma_{s8} = 8\pi^2 \eta_0^2 (\eta'_{s8} - 1)^2 \cdot b^v / \lambda^2$$

where η'_{s8} = index of refraction of the RBC hemoglobin at 800 nm relative to plasma η_0 (the plasma index of refraction), and,

$$S_{b13} = H (1-H) \sigma_{s13} \text{ (also from equation 7) where:}$$

$$\sigma_{s13} = 8\pi^2 \eta_0^2 (\eta'_{s13} - 1)^2 \cdot b^v / \lambda^2$$

η'_{s13} = the index of refraction of glucose at 1300 nm relative to η_0 .

Therefore, the 8 13 ratio has both hematocrit and glucose information. Whereas the $\alpha_1 \cdot \alpha'_1 / \Delta P$

(equation 18a) ratio has only hematocrit information. Therefore the differential combination of those ratios will be a strong function of glucose only.

E. Improved Accuracy Pulse Oximeter Device

The accuracy of present day pulse oximeters suffers from 4 major problems: tissue perfusion (low X_b and low $\partial X_b/\partial t$), d dependence (varying finger sizes), tissue nonhomogeneity (the tissue penetration depth for 660 nm light is not the same as for 940 nm light), and H dependence (see equation (5)).

All of the above mentioned deficiencies in pulse oximetry can be eliminated by understanding equation (13). Equation (13) indicates an "offset term", $-\frac{1}{A} \frac{\partial A}{\partial X_b}$.

Hence, while merely dividing $(\Delta i/i)_{\lambda 1}$ by $(\Delta i/i)_{\lambda 2}$ mitigates the effect of $\partial X_b/\partial t$, the d 's do not completely cancel, thereby yielding the above mentioned problems. To improve pulse oximeter accuracy, a derivative is needed as in (14), which eliminates the "offset term". Hence, the ratio of $(\partial \alpha/\partial t)_{805}/(\partial \alpha/\partial t)_{660}$ results in no H , d , or X_b dependence and the use of the multiple LED array and high-speed sampling as mentioned in section IV qualifies the tissue as homogeneous.

V. A Simplified Two Step Approach

(A) Determination of H and X_b

The bulk attenuation coefficient, α , can be easily measured with the optical array, at 805 nm, utilizing equation (10) and as described in Section IV(C). Notice that at 805 nm, α is a strong function of H and X_b since K_{s1} , K_{w1} , K_{p1} are small, see Figure 12.

Therefore, by knowing X_b itself, H can be determined. X_b itself can be determined using a

strain gage in the following two step approach. Step One, measure the strain gage resistance when the finger is made bloodless, by squeezing finger, such as with a stepper motor. Step Two, measure the strain gage resistance when the finger is blood filled, for example by suction.

Mathematically, at 805 nm and when K_s , K_p , K_w are small, equation (3) is approximated by:

$$\alpha^2 \approx 3KS \quad (36)$$

or

$$\alpha^2 \approx 3[K_b X_b][S_b X_b + S_s X_s] \quad (37)$$

Substituting (5) and (7) into (37) yields:

$$0 = 3 \left[H \frac{\sigma_s}{V} X_b \right] \left[H(1-H)(1.4-H) \frac{\sigma_s}{V} X_b + S_s X_s \right] - \alpha^2 \quad (38)$$

With X_b and α measured and known, and with the σ 's and S_s , X_s approximately constant, H can be solved with a quadratic formula or a polynomial fit.

The strain gage determination of X_b is as follows:

Let V_o = the volume of a bloodless finger. Let V_f = the volume of blood filled finger, and again considering the finger as a cylinder:

$$V_o = \pi r^2 z = V_i + V_w \quad (39)$$

$$V_f = \pi R^2 z = V_b + V_s + V_w \quad (40)$$

and

$$\frac{V_o}{V_f} = \left(\frac{r}{R} \right)^2 \quad (41)$$

From equation (20)

$$X_b = \frac{V_f - V_o}{V_f} = 1 - \frac{V_o}{V_f} \quad (42)$$

Substituting (41) into (42):

$$X_b = 1 - \left(\frac{r}{R}\right)^2 \quad (43)$$

Where the strain gage resistances are proportional to the radius, r and R , of the finger.

(B) Determination of tissue water content X_w

Choosing the wavelength of 1300 nm, where K_s and K_w are significant, the tissue water content, X_w , can be determined. Recall that $1 - X_b - X_w = X_s$ and substituting into (3) yields:

$$\begin{aligned} \alpha^2_{13} &= 3((K_b - K_s)X_b + (K_w - K_s)X_w + K_s) \\ &((\{K_b - K_s\} + \{S_b - S_s\})X_b + [\{K_w - K_s\} - S_s]X_w + (K_s + S_s)) \end{aligned} \quad (44)$$

With α_{13} , X_b and H determined and because K_b , K_s , K_w , S_b , and S_s are known coefficient values at 1300 nm, X_w is solved with either a quadratic formula or a polynomial fit.

RESULTS

Figure 13 demonstrates preliminary results with 30 patients the application of the method and apparatus and the application of Equation 19 on numerous patients with a correlation of $r = 0.96$.

As implied throughout, those skilled in the art will also appreciate that the methods for determining blood hematocrit values within the scope of the present invention may be adapted for determining other non-hematocrit biologic constituent values such as glucose, bilirubin, cholesterol, tissue water, etc.

The present invention may be embodied in other specific forms without departing from its spirit or essential characteristics. While the foregoing described embodiments are to be considered in all respects only as illustrative of the claimed invention, they are not intended to restrict the scope of the claims. The scope of the invention is, therefore, indicated by the following appended claims rather than by the foregoing description. All changes within the meaning and range of equivalency of the claims are to be embraced within their scope.

What is claimed is:

1. A method for determining a desired biologic constituent concentration of the blood of a patient, the blood flowing in a pulsatile fashion in a body part of the patient so as to be subjectable to transcutaneous examination in the body part, the body part defining a blood conduit and the method comprising the steps of:

- (a) placing the blood conduit within a blood conduit receiver with the blood flowing in the blood conduit;
- (b) directing radiation into the flowing blood within the blood conduit using a radiation generator situated within said blood conduit receiver means, said radiation defining a directed radiation comprising a first quantity of radiation at a radiation wavelength which, when directed into the flowing blood within the blood conduit,
 - (A) has a first attenuation value which varies with the desired biologic constituent concentration in the flowing blood and
 - (B) has a second attenuation value which varies with the concentration of components other than the desired biologic constituent in the flowing blood, which second attenuation value is at least ten times smaller than said first attenuation value, and
- (c) detecting the portion of said directed radiation which passes through both the blood conduit and the flowing blood therein using a radiation detector situated within said blood conduit receiver, said detected portion of said directed radiation comprising a second quantity of radiation at the radiation wave length; and

- (d) detecting energy from the flowing blood within the blood conduit using an energy transducer situated within said blood conduit receiver, said energy defining a transduced energy comprising a quantity of energy which when detected from the flowing blood within the blood conduit, has a value which varies with the normalized change of the pulsatile blood; and
- (e) operating exclusively on the second quantities of the radiation and the transduced energy to determine the desired biologic constituent concentration.

2. A method as defined in claim 1, wherein the step of detecting the second quantity of the radiation wavelength comprises the steps of:

- (a) determining the intensity of the total radiation wavelength; and
- (b) determining a radiation wavelength pulsatile value representing the intensities of a pulsatile component of the radiation wavelength at discrete time intervals during the pulse.

3. A method as defined in claim 1, wherein the step of detecting the transduced energy comprises the steps of:

- (a) determining the electronic signal generated from the transduced energy; and
- (b) determining a transduced energy pulsatile value representing the intensities of a pulsatile component of the transduced energy at discrete time intervals during the pulse.

4. A method as defined in claim 1, wherein the step of operating exclusively on the second quantities of the radiation at the radiation wavelength to determine the desired biologic constituent concentration of the patient comprises the steps of:

- (a) mathematically operating on the second quantities of the radiation wavelength such that the time derivative of the pulsatile intensities is normalized by the average intensity over the pulse interval followed by a distance derivative of that quantity to produce a value proportional to $\partial\alpha/\partial t$; and
- (b) mathematically operating on the second quantities of the radiation wavelength such that the logarithm of the intensity is distance differentiated to produce the value α .

5. A method as defined in claim 1, wherein the step of operating exclusively on the transduced energy comprises the step of performing the time derivative of the normalized pulsatile transduced energy to obtain the value $\partial X_b/\partial t$.

6. A method as defined in claim 1, wherein the step of operating exclusively on the second quantities of the radiation and the transduced energy comprises the step of mathematically solving the relationship $K_b = B \cdot (\alpha \cdot \partial\alpha/\partial t) / (\partial X_b/\partial t)$ with a polynomial function or empirically determined value.

7. A method as defined in claim 1, wherein the desired biologic constituent comprises hematocrit or hemoglobin.

8. A method as defined in claim 1, wherein the first attenuation value is substantially the same amount for oxyhemoglobin and for reduced hemoglobin in the flowing blood and the second attenuation value is at least ten items smaller than said first attenuation value for any competing constituent in the flowing blood.

9. A method as defined in claim 1, wherein the radiation wavelength is in the range from about 790 nanometers to 850 nanometers.

10. A method as defined in claim 1, wherein the radiation wavelength is in the range from about 550 nanometers to 600 nanometers.

11. A method as defined in claim 1, wherein the energy transducer means is a pressure transducer element, a strain gage element, a piezo electric film element, or a doppler detection element.

12. A method for determining a desired biologic constituent concentration of the blood of a patient, the blood flowing in a pulsatile fashion in a body part of the patient so as to be subjectable to transcutaneous examination in the body part, the body part defining a blood conduit and the method comprising the steps of:

- (a) placing the blood conduit within a blood conduit receiver with the blood flowing in the blood conduit;
- (b) directing radiation into the flowing blood within the blood conduit using a

radiation generator situated within said blood conduit receiver, said radiation defining a directed radiation comprising:

- (i) a first quantity of radiation at a radiation wavelength which, when directed into the flowing blood within the blood conduit,
 - (A) has a first attenuation value which varies with the desired biologic constituent concentration in the flowing blood and
 - (B) has a second attenuation value which varies with the concentration of components other than the desired biologic constituent in the flowing blood, which second attenuation value is at least ten times smaller than said first attenuation value, and
- (ii) a first quantity of a radiation at a second radiation wavelength, distinct from said first wavelength, which, when directed into the flowing blood within the blood conduit,
 - (A) has a third attenuation value which for varying concentrations in the flowing blood of the desired blood constituent is a non-fixed multiple of said first attenuation value; and
 - (B) has a fourth attenuation value which varies with the concentration of components other than the desired biologic constituent in the flowing blood, which fourth attenuation value is at least ten times greater than said second attenuation value;
- (c) detecting the portion of said directed radiation which passes through both the blood conduit and the flowing blood therein using a radiation detector situated within said

blood conduit receiver, said detected portion of said directed radiation comprising:

- (i) a second quantity of a radiation at the first radiation wavelength and,
 - (ii) a second quantity of a radiation at the second radiation wavelength;
- (d) detecting energy from the flowing blood within the blood conduit using an energy transducer situated within said blood conduit receiver, said energy defining a transduced energy comprising a quantity of energy which when detected from the flowing blood within the blood conduit, has a value which varies with the normalized change of the pulsatile blood; and
- (e) operating exclusively on the second quantities of the radiations and the transduced energy to determine the desired biologic constituent concentration.

13. A method as defined in claim 12, wherein the step of operating exclusively on the transduced energy comprises the step of performing the time derivative of the normal pulsatile transduced energy of the second radiation wavelength as defined in claim 19 to obtain the value $\partial X_2/\partial t$.

14. A method as defined in claim 12, wherein the step of operating exclusively on the second quantities of the radiation and the transduced energy comprises the step of solving the relationship $f(H) = G \cdot (\alpha \cdot (\partial\alpha/\partial t))$ first / $(\alpha \cdot (\partial\alpha/\partial t))$ second with a polynomial function or empirically determined value.

15. A method for determining a desired biologic constituent concentration of the blood

of a patient, the blood flowing in a pulsatile fashion in a body part of the patient so as to be subjectable to transcutaneous examination in the body part, the body part defining a blood conduit and the method comprising the steps of:

- (a) placing the blood conduit within a blood conduit receiver with the blood flowing in the blood conduit;
- (b) directing radiation into the flowing blood within the blood conduit using a radiation generator situated within said blood conduit receiver, said radiation defining a directed radiation comprising a first quantity of a radiation at a radiation wavelength which, when directed into the flowing blood within the blood conduit,
 - (A) has a first attenuation value which varies with the desired biologic constituent concentration in the flowing blood and
 - (B) has a second attenuation value which varies with the concentration of components other than the desired biologic constituent in the flowing blood, which second attenuation value is at least ten times smaller than said first attenuation value;
- (c) detecting the portion of said directed radiation which passes through both the blood conduit and the flowing blood therein using a radiation detector situated within said blood conduit receiver, said detected portion of said directed radiation comprising a second quantity of radiation at the radiation wavelength; and
- (d) detecting energy from the flowing blood within the blood conduit using an energy transducer situated within said blood conduit receiver, said energy defining a transduced energy comprising a quantity of energy which when detected from the

flowing blood within the blood conduit, has a value which varies with the normalized blood volume; and

- (e) operating exclusively on the second quantity of the radiation and the transduced energy to determine the desired biologic constituent concentration.

16. A method as defined in claim 15, wherein the step of operating exclusively on the transduced energy comprises the step of measuring the transduced energy when the blood conduit is blood-filled, then later made blood-less in order to obtain the value X_b .

17. A method as defined in claim 16, wherein the step of determining X_b is accomplished by solving $1 - (V_d/V_f)$ with the above energy transducer means.

18. A method as defined in claim 17, wherein the step of determining V_d/V_f is accomplished by solving $(V_c/V_f) - 1$ with a polynomial function of the pressure transducer element.

19. A method for determining a desired biologic constituent concentration of the blood of a patient, the blood flowing in a pulsatile fashion in a body part of the patient so as to be subjectable to transcutaneous examination in the body part, the body part defining a blood conduit and the method comprising the steps of:

- (a) placing the blood conduit within a blood conduit receiver with the blood flowing in the blood conduit;

- (b) directing radiation into the flowing blood within the blood conduit using a radiation generator situated within said blood conduit receiver, said radiation defining a directed radiation comprising:
 - (i) a first quantity of a radiation at a radiation wavelength which, when directed into the flowing blood within the blood conduit,
 - (A) has a fifth attenuation value which greatly varies with the undesired biologic constituent concentration in the flowing blood and
 - (B) has a sixth attenuation value which varies with the concentration of the desired biologic constituent in the flowing blood, which second attenuation value is at least ten times smaller than said fifth attenuation value, and
 - (ii) a first quantity of a radiation at a second radiation wavelength, distinct from said first wavelength, which, when directed into the flowing blood within the blood conduit,
 - (A) has a seventh attenuation value which for varying concentrations in the flowing blood of the undesired blood constituent is a multiple of said fifth attenuation value;
 - (B) has an eighth attenuation value which varies with the concentration of the desired biologic constituent in the flowing blood, which eighth attenuation value is at least five times greater than said sixth attenuation value;

- (c) detecting the portion of said directed radiation which passes through both the blood conduit and the flowing blood therein using a radiation detection means situated within said blood conduit receiving means, said detected portion of said directed radiation comprising:
 - (i) a second quantity of a radiation at the first radiation wavelength, and
 - (ii) a second quantity of a radiation at the second radiation wavelength;
- (d) detecting energy from the flowing blood within the blood conduit using an energy transducer means situated within said blood conduit receiving means, said energy defining a transduced energy comprising a quantity of energy which when detected from the flowing blood within the blood conduit, has a value which varies with the normalized change of the pulsatile blood; and
- (e) operating exclusively on the second quantities of the radiations and the transduced energy to determine the desired biologic constituent concentration.

20. A method for determining a desired biologic constituent concentration of the blood of a patient, the blood flowing in a pulsatile fashion in a body part of the patient so as to be subjectable to transcutaneous examination in the body part, the body part defining a blood conduit and the method comprising the steps of:

- (a) placing the blood conduit within a blood conduit receiver means with the blood flowing in the blood conduit;
- (b) directing radiation into the flowing blood within the blood conduit using a radiation generator situated within said blood conduit receiver, said radiation

defining a directed radiation comprising a first quantity of a radiation at a radiation wavelength which, when directed into the flowing blood within the blood conduit,

- (A) has a first attenuation value which varies with the desired biologic constituent concentration in the flowing blood and
- (B) has a second attenuation value which varies with the concentration of components other than the desired biologic constituent in the flowing blood, which second attenuation value is at least ten times smaller than said first attenuation value;
- (c) detecting the portion of said directed radiation which passes through both the blood conduit and the flowing blood therein using a radiation detector situated within said blood conduit receiver, said detected portion of said directed radiation comprising a second quantity of a radiation at the radiation wavelength; and
- (d) detecting energy from the flowing blood within the blood conduit using an energy transducer situated within said blood conduit receiver, said energy defining a transduced energy comprising a quantity of energy which when detected from the flowing blood within the blood conduit, has a value which varies with the normalized change of the pulsatile blood;
- (e) operating exclusively on the second quantities of the radiations and the transduced energy to determine the desired biologic constituent concentration by qualifying the tissue's homogeneity from the linearity of the distance differentiation.

21. A method as defined in claim 20, wherein the step of operating exclusively on the

second quantities of the radiation at the radiation wavelength to determine the desired biologic constituent concentration of the patient comprises the steps of:

- (a) mathematically operating on the second quantities of the radiation wavelength such that the time derivative of the pulsatile intensities is normalized by the average intensity over the pulse interval followed by a distance derivative of that quantity to produce a value proportional to $\partial\alpha/\partial t$;
- (b) mathematically operating on the second quantities of the radiation wavelength such that the logarithm of the intensity is distance differentiated to produce the value α ;
- (c) mathematically determining the linearity and deviation of the logarithm of the intensity and the $(\partial I/\partial t)/I$ values versus distance; and
- (d) mathematically decoupling, isolating, and determining the individual constituent absorptive and scattering coefficients from the homogeneity qualified α , $\partial\alpha/\partial t$ and $\partial X_s/\partial t$ values.




FIG. 1




FIG. 1A-transmission




FIG. 1B - reflectance




FIG. 1C - reflectance




FIG 1D - mylar array






FIG. 2





6-4-47





FIG. 4

FIG. 5
 $(V_c / V_l - 1) = 1.353 + 1128 / P$






FIG. 6

FIG. 7

FIG. 7






FIG. 8








FIG. 12





FIG. 13



INTERNATIONAL SEARCH REPORT

International application No.
PCT/US99/02586

A. CLASSIFICATION OF SUBJECT MATTER

IPC(6) : A61B 5/00

US CL : 600/322

According to International Patent Classification (IPC) or to both national classification and IPC

B. FIELDS SEARCHED

Minimum documentation searched (classification system followed by classification symbols)

U.S. : 600/310, 322, 326, 328

Documentation searched other than minimum documentation to the extent that such documents are included in the fields searched

Electronic data base consulted during the international search (name of data base and, where practicable, search terms used)

C. DOCUMENTS CONSIDERED TO BE RELEVANT

Category*	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
X	US 5,372,136 A (STEUER et al) 13 December 1994, entire document.	1-3, 7-10, 12, 15, 19, 20
A	US 5,553,615 A (CARIM et al) 10 September 1996, entire document.	1-21

☐ Further documents are listed in the continuation of Box C. ☐ See patent family annex.

* Special categories of cited documents:	*T* later document published after the international filing date or priority date and not in conflict with the application but cited to understand the principle or theory underlying the invention
A document defining the general state of the art which is not considered to be of particular relevance	*X* document of particular relevance; the claimed invention cannot be considered novel or cannot be considered to involve an inventive step when the document is taken alone
E earlier document published on or after the international filing date	*Y* document of particular relevance; the claimed invention cannot be considered to involve an inventive step when the document is combined with one or more other such documents, such combination being obvious to a person skilled in the art
L document which may throw doubts on priority claim(s) or which is cited to establish the publication date of another citation or other special reason (as specified)	*A* document member of the same patent family
O document referring to an oral disclosure, use, exhibition or other means	
P document published prior to the international filing date but later than the priority date claimed	

Date of the actual completion of the international search

08 APRIL 1999

Date of mailing of the international search report

14 MAY 1999

Name and mailing address of the ISA/US
Commissioner of Patents and Trademarks
Box PCT
Washington, D.C. 20231

Facsimile No. (703) 305-3230

Authorized officer

ERIC F. WINAKUR

Telephone No. (703) 308-3940

**This Page is Inserted by IFW Indexing and Scanning
Operations and is not part of the Official Record**

BEST AVAILABLE IMAGES

Defective images within this document are accurate representations of the original documents submitted by the applicant.

Defects in the images include but are not limited to the items checked:

- ☐ **BLACK BORDERS**
- ☐ **IMAGE CUT OFF AT TOP, BOTTOM OR SIDES**
- ☒ **FADED TEXT OR DRAWING**
- ☐ **BLURRED OR ILLEGIBLE TEXT OR DRAWING**
- ☐ **SKEWED/SLANTED IMAGES**
- ☐ **COLOR OR BLACK AND WHITE PHOTOGRAPHS**
- ☐ **GRAY SCALE DOCUMENTS**
- ☐ **LINES OR MARKS ON ORIGINAL DOCUMENT**
- ☐ **REFERENCE(S) OR EXHIBIT(S) SUBMITTED ARE POOR QUALITY**
- ☐ **OTHER:** _____

IMAGES ARE BEST AVAILABLE COPY.

As rescanning these documents will not correct the image problems checked, please do not report these problems to the IFW Image Problem Mailbox.